

# The Engineering Guide to Oxygen Measurement

Applications from Fuel Cells to Natural Gas



---

*Comprehensive, technology-neutral handbook*  
**Includes principles, selection guidance, safety integrity (SIL)  
considerations, and illustrated schematics**

---

***Modcon Systems Ltd. | Engineering Handbook Edition @ 2025 | Rev.2.0***

## Contents

1.	Measurement Fundamentals and Terminology .....	3
2.	System View: Analyzer + Sampling + Installation .....	5
3.	Gas-Phase Oxygen Measurement Technologies .....	9
3.1	Electrochemical (Galvanic / Polarographic) – Gas Phase .....	9
3.2	Zirconia (Solid Electrolyte / Nernst Cell) .....	11
3.3	Paramagnetic (Dumbbell / Magneto-pneumatic / Thermomagnetic).....	13
3.4	Laser Absorption (TDLAS).....	15
3.5	Optical Luminescence / Fluorescence Quenching (Gas).....	17
3.6	Gas Chromatography (GC) and Multi-Component Methods .....	19
3.7	Mass Spectrometry (MS) for Multi-Gas Analysis .....	19
3.8	Thermal Conductivity (TCD) as an Indirect Approach .....	20
3.9	Wet-chemistry / Laboratory Reference Methods.....	21
4.	Dissolved Oxygen (DO) Measurement Technologies .....	22
4.1	Clark-type Electrochemical DO Sensors .....	22
4.2	Optical DO (Luminescence Lifetime / Phase).....	23
5.	Comparative Decision Tables .....	24
6.	Safety Integrity (SIL) and Failure Modes .....	25
7.	Commissioning, Calibration and Diagnostics .....	26
8.	MOD-1040 Optical Oxygen Analyzer .....	29
9.	Applications of Oxygen Measurement in Industrial Processes .....	35
9.1	Natural Gas Processing and Transmission .....	35
9.2	Hydrogen Production, Processing, and Storage .....	36
9.3	Chemical and Petrochemical Processing .....	37
9.4	High-Pressure Gas Production and Processing .....	38
9.5	Cross-Application Considerations .....	38
10.	Glossary and Terms .....	39

## 1. Measurement Fundamentals and Terminology

Oxygen is a chemical element with the symbol O and atomic number 8; in industrial measurement contexts it is most commonly encountered as the diatomic molecule  $O_2$ , a colorless, odorless, and tasteless gas at standard ambient conditions. It is a highly reactive non-metal and a potent oxidizing agent, readily forming compounds (oxides) with nearly all elements except noble gases. Molecular oxygen constitutes about 21 % of Earth's atmosphere by volume and is the second most abundant atmospheric gas, making it a central measurand in both environmental and process industries.




Fig. 1 – Oxygen Measurement Fundamentals

### Physical and Chemical Basis for Measurement

#### Forms and States:

Under typical process conditions, oxygen exists as a stable **diatomic gas ( $O_2$ )**, although other forms (allotropes such as ozone  $O_3$  or transient species like singlet oxygen) exist under special conditions. Molecular oxygen can condense into a pale blue liquid at cryogenic temperatures ( $-183\text{ }^\circ\text{C}$ ) and freeze at  $-218\text{ }^\circ\text{C}$ ; while these states are rarely relevant to in-process measurement, they underscore the range of physical behavior oxygen exhibits.

#### Paramagnetism:

A unique physical property of dioxygen in its ground state is paramagnetism—an attraction toward magnetic fields due to unpaired electrons. This property underpins the operation of paramagnetic oxygen analyzers and reinforces the importance of linking molecular behavior to measurement modality.

**Solubility and Phase Interaction:**

Oxygen dissolves more readily in water than nitrogen, and its solubility is temperature-dependent; cold water contains significantly more dissolved oxygen than warm. At 25 °C and 1 atm, freshwater holds ~6 mL O<sub>2</sub> per liter, increasing to ~9 mL at 5 °C. This relationship is critical for defining dissolved oxygen (DO) measurement ranges and calibration strategies.

**Measurand:**

The measurand defines *what* is being measured—not all oxygen measurement is simply “percent O<sub>2</sub>.” Depending on industry and system purpose, oxygen may be specified as:

- Volume fraction (% v/v or ppmv) — common in gas streams and safety analyses.
- Partial pressure (pO<sub>2</sub>) — relevant in high-pressure or controlled-atmosphere systems.
- Dissolved oxygen (mg/L or ppb) — central in aqueous corrosion, boiler feedwater, and bioprocess applications.

Each form requires careful unit consistency in calibration and reporting.

**Partial Pressure vs. Volume Fraction:**

Under process conditions, the partial pressure of oxygen will vary with total pressure and temperature according to Dalton’s Law of Partial Pressures. Thus, specifying only % v/v is insufficient unless accompanied by a defined pressure/temperature reference. This distinction is fundamental in combustion, high-pressure hydrogen, and vacuum systems.

**Range and Resolution:**

Engineers must specify both the measurement range (maximum–minimum oxygen concentration of interest) and the required resolution/uncertainty within that range. Trace applications, such as hydrogen safety or inerting gas verification, often require resolution well into the ppmv range; by contrast, combustion control may focus on percent-level oxygen.

**Response Time:**

Response time (e.g., *T*<sub>90</sub>, time to reach 90 % of a step change) reflects both sensor physics and system transport dynamics. In-situ measurements typically have faster true response than extractive systems that include sampling lines, filters, and pressure reduction stages.

**Air Ingress and Adsorption/Desorption:**

A central practical limitation in trace oxygen measurement is the influence of air ingress through leaks or permeation, and adsorption/desorption effects on wetted surfaces. These system phenomena—not sensor physics—often dominate achievable uncertainty in low-ppmv applications.

## 2. System View: Analyzer + Sampling + Installation

Oxygen analyzers must be evaluated as complete systems. Extractive systems add transport delay and introduce leak points; in-situ systems reduce delay but require robust process interfaces and fouling control. For trace oxygen, sample system design often dominates both bias and response time.




Fig. 2 – Extractive sampling system vs In-situ System

Oxygen analysis in industrial applications is seldom limited to the physical sensing element alone. Instead, the complete measurement system—including sampling, transport, conditioning, installation configuration, and analyzer diagnostics—largely determines performance. An analyzer’s nominal specification (range, accuracy, resolution) represents the sensor’s intrinsic capability, but it is the system architecture that governs real-world bias, response time, susceptibility to contamination, and overall reliability.

In any practical implementation, the measurement loop consists of four interdependent domains:

1. Process Interface and Installation
2. Sampling or In-Situ Interaction
3. Sample Conditioning and Transport
4. Analyzer Physics, Calibration, and Diagnostics

Each domain contributes to measurement error, uncertainty, or risk if not engineered appropriately.

### 2.1 Process Interface and Installation

The choice of installation point and interface directly influences whether the measurement reflects the true process state. Key architectural considerations include:

- **Representative Location:** An analyzer must sample where the gas composition truly reflects process conditions (e.g., well-mixed flow, away from boundary layers or recirculation regions). Poor placement can lead to persistent bias and delayed detection of excursions.
- **Mechanical Integrity:** Sample probes, process connections, and seals must be designed to maintain leak-tightness under all service conditions. Even microleaks introduce ambient air, which at trace oxygen levels rapidly overwhelms the measurand and invalidates readings.
- **Environmental Exposure:** Direct exposure to sunlight, precipitation, temperature swings, or mechanical vibration affects instruments differently. For extractive analyzers, outdoor installations often require sun shields, insulated sample lines, and environmental enclosures. In contrast, in-situ devices may need thermal isolation or purge barriers to avoid condensation or fouling.
- **Hazardous Areas:** Oxygen measurement commonly intersects with explosive atmospheres (e.g., hydrogen production, natural gas streams). All hardware must comply with applicable hazardous-area classifications and protective methods (e.g., ATEX, IECEx, NEC), and installation must ensure that no ignition source exists near the measured gas stream.

## 2.2 Extractive vs. In-Situ Sampling

The system approach diverges fundamentally depending on whether the analyzer is extractive or in-situ.

Extractive sampling draws process gas through a probe and a series of conditioning elements to the analyzer located remotely. While common, this approach introduces transport delay, adsorption/desorption phenomena, condensation risk, and multiple potential leak points:

- **Transport Delay:** The time required to move gas from point of interest to the analyzer extends response time beyond the intrinsic sensor dynamics. In safety applications (e.g., purge monitoring), excessive delay can mean a hazardous condition is not observed until after an unsafe state has existed.
- **Adsorption/Desorption:** Gas molecules—including oxygen—can adhere to the internal walls of sample tubing, regulators, and fittings. These surface interactions cause **memory effects**, where the measured oxygen signal reflects both current and historical concentrations.
- **Pressure and Temperature Variation:** Pressure regulators and temperature gradients between process and analyzer environments complicate not only the sample path but can introduce condensation, which affects many sensor types (e.g., paramagnetic and electrochemical cells) unless properly heated or dried.

Because of these issues, extractive systems often require elaborate conditioning trains (filters, heaters, dryers, pressure regulators) that must be carefully engineered, monitored,



and maintained. In some modern systems, components such as heated sample lines and in-line dewpoint control are specified to ensure the sample remains in a consistent, measurand-representative state.

### In-Situ Sensing

In-situ analyzers, by contrast, mount the sensing element directly into the process pipe or duct. This architecture eliminates transport delay and most sample transport errors, but places a premium on:

- **Robust Process Interfaces:** The probe or sensor module must withstand process temperature, pressure, and particulate or corrosive environments while maintaining optical or electrochemical integrity.
- **Fouling and Contamination Control:** Direct exposure can accelerate degradation of optical windows or sensor membranes. System design must facilitate easy cleaning, effective purge strategies, or self-monitoring diagnostics that identify contamination before it compromises measurement.

In modern industrial analyzers using optical methods such as Tunable Diode Laser Absorption Spectroscopy (TDLAS)—which performs measurements without direct gas–component contact—the in-situ approach synergizes with the physics of the technology to minimize maintenance and maximize response speed. These systems also incorporate advanced diagnostics (e.g., laser line-lock checks) to continuously validate measurement integrity.

### 2.3 Sample Conditioning and Its Impacts

Whether extractive or semi-extractive, sample conditioning plays a crucial role in preserving the measurand’s representativeness:

- **Filtration:** Removes particulates but can also trap or alter oxygen content if not properly sized or cleaned.
- **Pressure Regulation:** Necessary to bring sample to analyzer’s operating range, but regulators add dead volume and potential for oxygen ingress or pressure-induced adsorption effects.
- **Moisture Management:** Condensation downstream of cool spots biases many sensor types. Heated sample lines or active moisture control are common solutions, particularly when water vapor concentrations are high.

In some advanced spectroscopic systems, non-contact optical paths and heated sample cells allow measurement of “hot, wet” gases without complex dehydration steps, offering superior fidelity in challenging environments.

### 2.4 Analyzer Calibration, Diagnostics, and System Diagnostics

The analyzer itself must be engineered with **internal diagnostics** and calibration strategies that complement the installation and sampling approach:



- **Internal Reference Cells:** Techniques such as TDLAS often incorporate sealed reference cells to maintain “line-lock” on the desired absorption wavelength, enabling continuous validation of calibration without frequent external gas checks.
- **Span and Zero Verification:** Even analyzers factory-calibrated to traceable standards should support field checks using known reference gases. For extractive systems, span and zero checks help account for integrity of the sample path; for in-situ systems, they help isolate drifts in optical sources or detectors.
- **Self-Check Diagnostics:** Modern analyzers monitor internal conditions (light source stability, signal strength, detector noise) and can flag anomalies indicative of contamination, optical alignment drift, or component degradation.
- **Proof-Testing in Safety Applications:** For Safety Instrumented Functions (SIFs), calibration and diagnostics must be part of a formal proof test strategy, verifying that dangerous failures are detected within defined intervals and that the measurement loop responds to unsafe conditions within required times.

### 3.5 System Behavior and Measurement Uncertainty

A key insight from industrial practice is that system uncertainty is rarely determined solely by sensor specification. Instead, the combined effects of sampling transport and conditioning, process interface design, analyzability under dynamic conditions, and environmental exposures define the practical performance envelope. For high-integrity and safety applications, engineers must consider:

- Total system response time
- Bias introduced by surface interactions
- Leak susceptibility at interfaces
- Diagnostic coverage and failure modes
- Installation environment and accessibility for maintenance

By evaluating the complete measurement system instead of the sensor alone, designers can choose technologies and configurations that minimize system-level error sources and ensure measurement reliability across operating conditions.

### 3. Gas-Phase Oxygen Measurement Technologies

#### 3.1 Electrochemical (Galvanic / Polarographic) – Gas Phase

Electrochemical oxygen sensors measure oxygen through reduction reactions at a cathode. Oxygen reaches the electrode by diffusion through a membrane or capillary barrier, making the signal fundamentally diffusion-limited.

Galvanic sensors are self-powered cells where oxygen participates in the electrochemistry. Polarographic sensors use an external bias voltage to control the cathode reaction. Both require stable membrane condition and compensation for temperature and pressure.

These sensors can be sensitive at low ppm levels in clean gases and are widely used in portable and low-cost fixed installations. The primary drawbacks are consumable life, drift, and susceptibility to poisoning from reactive gases and humidity.

In trace oxygen service, engineering success often depends more on leak integrity and sample design than on sensor resolution.

#### Electrochemical Oxygen Sensor




Fig. 3 – Electrochemical oxygen sensor (gas) – diffusion-limited cell

#### *Safety integrity and SIL considerations*

Electrochemical sensors generally provide limited intrinsic diagnostics for drift and poisoning. For safety-related use, apply redundancy (e.g., 1oo2), tight proof-test intervals, and conservative trip points. Treat the sample system as part of the SIF because air ingress and flow failures can dominate output.

*Failure-mode-focused SWOT (with diagnostic coverage)*

<b>Dimension</b>	<b>Key points</b>	<b>Typical fail-high behavior</b>	<b>Typical fail-low behavior</b>
<b>Strengths</b>	Low initial cost; compact; strong sensitivity in clean service; simple electronics; broad availability.		
<b>Weaknesses</b>	Consumable cell and electrolyte; drift/aging; sensitive to humidity, temperature, pressure/flow; poisoning risk.	Air ingress upstream; membrane damage increasing diffusion; calibration gas contamination.	Electrolyte depletion; cathode poisoning; membrane blockage; insufficient flow; flooding by moisture.
<b>Opportunities</b>	Portable verification; secondary measurements; applications with frequent maintenance access.		
<b>Threats</b>	Undetected drift between checks; air ingress causes false high oxygen; poisoning causes false low; high lifecycle cost in harsh duty.	Air ingress upstream; membrane damage increasing diffusion; calibration gas contamination.	Electrolyte depletion; cathode poisoning; membrane blockage; insufficient flow; flooding by moisture.
<b>Diagnostics / Proof testing</b>	Typically limited. Improve diagnostic coverage with flow/pressure supervision, as-found/as-left trending, bump tests, and comparison to a diverse technology.		

### 3.2 Zirconia (Solid Electrolyte / Nernst Cell)

Zirconia oxygen sensors operate as solid-electrolyte concentration cells at elevated temperature. With platinum electrodes and a reference gas, they generate a voltage governed by the logarithm of oxygen partial pressure ratio (Nernst behavior).

The technology is dominant in combustion and flue gas applications due to fast response at percent-level oxygen and compatibility with hot process environments.

Key sensitivities include thermal shock, soot/particulate deposition, condensation, and exposure to strongly reducing atmospheres. Heater control and reference-side integrity are central to stable measurement.

Applied within its intended domain with proper protection and diagnostics, zirconia provides robust control feedback for combustion efficiency and emissions strategies.

#### Zirconia (Nernst Cell) Oxygen Sensor




Fig. 4 – Zirconia (Solid Electrolyte / Nernst Cell) oxygen sensor

#### Safety integrity and SIL considerations

For safety-related combustion functions, integrate heater monitoring, temperature supervision, and plausibility checks. Consider redundancy when oxygen measurement prevents unsafe combustion conditions. Proof testing should validate reference integrity and response.

#### Failure-mode-focused SWOT (with diagnostic coverage)

Dimension	Key points	Typical fail-high behavior	Typical fail-low behavior
Strengths	Fast response in hot gases; mature technology; strong		

	combustion compatibility; direct %O <sub>2</sub> control variable.		
<b>Weaknesses</b>	Requires high-temperature operation; susceptible to condensation/thermal shock; electrode aging; protection required.	Reference path leak/contamination; temperature control error causing bias.	Heater failure; cracked electrolyte; condensation; electrode poisoning.
<b>Opportunities</b>	Boiler and furnace optimization; burner management support; emissions monitoring contexts.		
<b>Threats</b>	Heater failure; reference contamination; fouling leading to sluggish response; misuse outside intended range.	Reference path leak/contamination; temperature control error causing bias.	Heater failure; cracked electrolyte; condensation; electrode poisoning.
<b>Diagnostics / Proof testing</b>	Moderate: heater current/temperature, impedance checks, EMF plausibility. Proof tests should include step response and reference-side checks.		

### 3.3 Paramagnetic (Dumbbell / Magneto-pneumatic / Thermomagnetic)

Paramagnetic analyzers use oxygen's attraction to magnetic fields. Implementations convert magnetic-force effects into a mechanical deflection or pressure signal proportional to oxygen concentration.

They are widely used for percent-level measurement in clean, dry gas streams such as oxygen enrichment, air separation, and inerting.

Practical performance depends strongly on stable sample conditioning: moisture, dust, condensables, and pressure/flow instability can introduce bias. In many plants, the sample system determines reliability.

For inerting, verify that the measurement location and sample transport reflect the hazardous condition within required time.




Fig. 5 – Paramagnetic oxygen analyzer – magnetic susceptibility principle

#### *Safety integrity and SIL considerations*

Paramagnetic analyzers can support safety functions in clean services when sample conditioning is robust and monitored. Intrinsic diagnostics can be limited; improve safety integrity with conditioning alarms, flow/pressure monitoring, and periodic gas checks.

#### *Failure-mode-focused SWOT (with diagnostic coverage)*

Dimension	Key points	Typical fail-high behavior	Typical fail-low behavior
<b>Strengths</b>	High selectivity to O <sub>2</sub> ; non-consumable; stable in clean service; fast response.		
<b>Weaknesses</b>	Sensitive to pressure/flow and	Pressure regulation instability;	Sample dilution; moisture ingress;

	contamination; some designs sensitive to vibration/temperature gradients.	mechanical imbalance; temperature gradient bias.	blocked filters reducing flow; downstream leaks.
<b>Opportunities</b>	ASU and enrichment; clean-gas inerting; long-term monitoring where conditioning is reliable.		
<b>Threats</b>	Conditioning failures; dilution/leaks; installation-induced vibration and thermal gradients.	Pressure regulation instability; mechanical imbalance; temperature gradient bias.	Sample dilution; moisture ingress; blocked filters reducing flow; downstream leaks.
<b>Diagnostics / Proof testing</b>	Often modest. Improve DC with conditioning instrumentation (dew point, flow, pressure), plausibility checks, and scheduled proof tests.		

### 3.4 Laser Absorption (TDLAS)

TDLAS measures oxygen by tuning a diode laser across an oxygen absorption feature and interpreting the attenuation using absorption physics. The method can be applied in-situ or in extractive cells.

TDLAS is valued for very fast response and the ability to reduce sample-system risks by measuring directly in the process. It avoids consumable sensing elements.

Engineering challenges include window fouling, optical alignment, beam steering, and signal-to-noise in dusty or wet environments. These are mitigated by mechanical design, purge strategies, and using optical health diagnostics as primary protection layers.

For fast transients such as purge and inerting upsets, in-situ TDLAS can materially improve time-to-detection when the optical path is representative.




Fig. 6 – TDLAS oxygen measurement – laser absorption path

#### *Safety integrity and SIL considerations*

TDLAS can provide high diagnostic coverage via signal strength and spectral fit metrics. In SIL use, treat optical degradation and out-of-model conditions as safety-relevant alarms/trips and define proof tests that validate both measurement and diagnostics.

#### *Failure-mode-focused SWOT (with diagnostic coverage)*

Dimension	Key points	Typical fail-high behavior	Typical fail-low behavior
<b>Strengths</b>	Very fast response; no consumables; in-situ capability; strong selectivity; rich diagnostics.		
<b>Weaknesses</b>	Requires optical access and	Baseline fitting error; detector	Window fouling; beam blockage;

	alignment; window condition critical; higher engineering effort.	saturation; stray light artifacts.	reduced signal; path length change unnoticed.
<b>Opportunities</b>	Fast safety detection; harsh monitoring where extractive sampling is risky; modernization of inerting systems.		
<b>Threats</b>	Fouling/condensation blocking optics; misalignment after maintenance; ignoring diagnostics.	Baseline fitting error; detector saturation; stray light artifacts.	Window fouling; beam blockage; reduced signal; path length change unnoticed.
<b>Diagnostics / Proof testing</b>	Typically high. Proof tests should validate diagnostic thresholds and safe-state response to optical faults.		

### 3.5 Optical Luminescence / Fluorescence Quenching (Gas)

Optical luminescence oxygen measurement relies on quenching of a luminophore. Oxygen reduces the luminescence intensity and/or lifetime. Lifetime/phase techniques reduce sensitivity to optical intensity drift.

Degradation mechanisms are typically linked to optical surfaces, coating integrity, and compensation model validity rather than chemical consumption. This can reduce routine maintenance in applications where electrochemical sensors degrade quickly.

Critical design elements include temperature and pressure compensation, contamination control, and stable optical coupling. Because signal quality can be monitored, intrinsic diagnostics can be strong.

In safety integrity contexts, the analyzer must be specified with clear behavior upon diagnostic failure and with proof tests that validate both measurement and diagnostic action.




Fig. 7 – Optical luminescence oxygen sensor – quenching / lifetime detection

#### *Safety integrity and SIL considerations*

Lifetime-based approaches can offer strong diagnostic coverage (signal quality, lifetime plausibility, sensor temperature checks). Define safe-state behavior for out-of-range optical metrics and use proof tests to confirm that diagnostics trigger the defined safety response.


#### *Failure-mode-focused SWOT (with diagnostic coverage)*

Dimension	Key points	Typical fail-high behavior	Typical fail-low behavior
Strengths	Non-consumable principle; strong stability with		

	lifetime methods; reduced flow dependence; strong diagnostics potential.		
<b>Weaknesses</b>	Requires T/P compensation; optical surfaces must remain clean; coating aging can change sensitivity.	Temperature miscompensation; phase measurement bias; optical reflections changing response.	Severe fouling; coating degradation; moisture films attenuating emission.
<b>Opportunities</b>	High-integrity hydrogen and inerting; long unattended operation; reduced sample-system complexity.		
<b>Threats</b>	Fouling/coating damage; miscompensation outside modeled envelope; poor installation thermal gradients.	Temperature miscompensation; phase measurement bias; optical reflections changing response.	Severe fouling; coating degradation; moisture films attenuating emission.
<b>Diagnostics / Proof testing</b>	High potential: lifetime plausibility, modulation depth, SNR. Proof tests should validate diagnostic thresholds and representative gas checks.		

### 3.6 Gas Chromatography (GC) and Multi-Component Methods


GC measures oxygen as a component of a multi-gas analysis. Sample injection and separation enable selective quantification, but cycle time is typically minutes. GC is best for specification, auditing, and multi-component quality control rather than fast safety trips.



*Fig. 8 – Gas chromatography overview – separation + detector*

### 3.7 Mass Spectrometry (MS) for Multi-Gas Analysis

MS provides rapid multi-gas capability by measuring a mass spectrum. It is powerful for diagnostics and optimization, but requires vacuum/inlet integrity and robust calibration.



*Fig. 9 – Mass spectrometry overview – ionization + m/z detection*

### 3.8 Thermal Conductivity (TCD) as an Indirect Approach

TCD infers oxygen via bulk thermal conductivity changes and is not inherently selective. It is most appropriate for stable, binary-like mixtures and is generally not recommended for safety functions in variable matrices.

#### Thermal Conductivity Detector





Fig. 10 – Thermal conductivity (TCD) – indirect measurement concept

### 3.9 Wet-chemistry / Laboratory Reference Methods

Laboratory reference methods provide validation and calibration support. They are not continuous process tools but can be essential in quality assurance and in verifying online analyzer drift.

#### Wet Chemistry / Titration



Titration → Endpoint

Fig. 11 – Wet-chemistry titration – reference concept

## 4. Dissolved Oxygen (DO) Measurement Technologies

### 4.1 Clark-type Electrochemical DO Sensors

Clark-type DO sensors are amperometric devices where dissolved oxygen diffuses through a membrane and is reduced at a cathode. They consume oxygen, so measurement can depend on flow and boundary layer thickness.




Fig. 12 – Clark-type DO sensor – membrane + electrochemistry

## 4.2 Optical DO (Luminescence Lifetime / Phase)

Optical DO uses luminescence quenching on an optode. Because oxygen is not consumed, optical DO reduces flow dependence and can provide improved stability for long deployments. Lifetime/phase methods improve robustness.




Fig. 13 – Optical DO sensor – optode + reader

## 5. Comparative Decision Tables

The tables below summarize practical selection criteria. They should be read together with the detailed technology sections. In safety-related functions, evaluate the complete loop: sensor + sampling/installation + transmitter/logic + proof testing and diagnostics.

Technology	Typical range	Best for	Response	Cross-sensitivities	Maintenance	Common pitfalls	SIL/Diagnostics
Electrochemical (gas)	ppm to %	Portable / low-cost, clean gases	s–min	Poisoning, humidity, pressure/flow	Consumable sensor	Air ingress dominates trace O <sub>2</sub>	Low intrinsic; needs proof tests
Zirconia	% (and some ppm)	Combustion/flue gas, high T	fast	Reducing gases, condensation	Heater, cell aging	Thermal shock, soot/particulates	Moderate via heater checks
Paramagnetic	%	Clean dry gases, enrichment	fast	Pressure/flow, vibration	Low/medium	Wet/dirty sample causes bias	Limited intrinsic; rely on stable sample
TDLAS	ppm to % (path dependent)	In-situ fast control, harsh conditions if optics protected	very fast	Spectral interference, optics	Low/medium	Window fouling and alignment	High diagnostics (signal health)
Optical luminescence (gas)	ppm to %	High integrity, low maintenance	fast	T/P compensation, fouling	Low	Optical surface contamination	Strong via lifetime + self-checks
GC	ppm to %	Multi-component, custody/spec	minutes	Peak overlaps, sampling	High	Transport delay, calibration	QA via chromatography metrics
MS	ppm to %	Multi-gas fast survey	seconds	Matrix effects, drift	High	Vacuum/inlet issues	Diagnostics via internal checks
TCD (indirect)	depends	Binary-like mixtures	fast	Non-selective	Low	Composition changes	Not recommended for SIF

## 6. Safety Integrity (SIL) and Failure Modes

Safety integrity assessment evaluates dangerous failure rates, diagnostic coverage, proof-test intervals, and common-cause failures across redundant channels. Oxygen analyzers used in safety functions should define safe behavior under internal fault and provide diagnostics wired into safety logic.

Fail-high and fail-low mapping depends on the SIF: in inerting, fail-low can be dangerous (missed oxygen ingress), whereas fail-high can cause nuisance trips. Therefore, SIL suitability is application-specific and must be evaluated per safety requirement specification.



Fig. 14 – Safety Integrity Level

## 7. Commissioning, Calibration and Diagnostics

Commissioning, calibration, proof testing, and diagnostics form the operational backbone of reliable oxygen measurement. Regardless of sensing technology, an oxygen analyzer must be treated as a measurement system whose performance is verified not only at the factory, but throughout its service life. These activities ensure that the analyzer delivers traceable, repeatable, and safety-relevant information under real process conditions.



Fig. 15 – Commissioning Checklist

### 7.1 Commissioning

Commissioning is the initial verification that the analyzer system has been correctly installed and is fit for service. This phase confirms that mechanical, electrical, and process interfaces meet design intent before routine operation begins.

Key commissioning activities include:

- Verification of correct installation location and process orientation to ensure representative sampling.
- Leak integrity testing of all process connections, sample lines, and fittings, particularly critical for trace oxygen applications where ambient air ingress can dominate the measured signal.
- Confirmation of operating pressure, temperature, and flow conditions against analyzer specifications.

- Validation of purge, vent, and depressurization philosophy to avoid trapped volumes that could accumulate oxygen.
- Electrical checks including power quality, grounding, signal wiring, and communication interfaces.

Commissioning establishes the baseline condition of the analyzer and provides reference documentation for future performance comparisons.

## 7.2 Calibration

Calibration aligns the analyzer output with known reference values of oxygen concentration. While many modern analyzers are factory-calibrated using traceable standards, field calibration or verification remains essential to confirm system-level performance, especially when sampling systems or process interfaces are involved.

Calibration procedures typically involve:

- Zero checks, using oxygen-free or near-zero reference gases (e.g., nitrogen with certified impurity limits).
- Span checks, using certified calibration gases at one or more points within the operating range.
- Documentation of as-found and as-left results to track drift over time.

Calibration gases and procedures should conform to recognized standards, including relevant ASTM practices for gas analysis and instrument validation, to ensure traceability and comparability of results. Calibration should not be viewed as a substitute for system integrity; repeated calibration adjustments may indicate underlying issues such as leaks, contamination, or sensor degradation.

## 7.3 Validation and Standards Compliance

Analyzer validation confirms that the instrument performs according to its intended measurement function within defined uncertainty limits. For oxygen analyzers, validation should follow established standards to ensure technical rigor and regulatory acceptance.

ASTM standards provide widely recognized methodologies for:

- Analyzer performance verification
- Accuracy and precision assessment
- Repeatability and response time evaluation
- Bias identification and correction

Adhering to ASTM standards during validation ensures that measurement results are technically defensible, reproducible across installations, and suitable for use in safety, quality, and regulatory contexts. Validation is particularly important when analyzers are used in critical applications such as inerting, corrosion prevention, or safety-instrumented functions.



## 7.4 Proof Testing

Proof testing is a structured, periodic activity intended to detect dangerous undetected failures, particularly for analyzers used in safety-related applications. Proof testing goes beyond calibration by explicitly verifying that the analyzer and its diagnostics can detect conditions that would otherwise compromise safety.

Typical proof testing may include:

- Simulation of known oxygen concentrations or process deviations
- Verification of alarm thresholds and signal transmission to control or safety systems
- Confirmation of analyzer response time under defined conditions
- Assessment of failure behavior (fail-high, fail-low, or fail-safe)

The proof-test interval should be defined based on failure modes, diagnostic coverage, and application risk—not solely on calibration frequency. For safety-critical services, proof testing is an essential element of functional safety lifecycle management.

## 7.5 Diagnostics and Continuous Health Monitoring

Modern oxygen analyzers increasingly incorporate internal diagnostics that continuously monitor sensor health and system integrity. Diagnostics may include checks on signal strength, sensor impedance, optical path integrity, heater function, or internal reference stability, depending on the measurement technology.

Effective diagnostics serve several purposes:

- Early detection of degradation before measurement accuracy is compromised
- Reduction of unnecessary maintenance or recalibration
- Support for condition-based maintenance strategies
- Improved confidence in measurement availability and reliability

Diagnostic outputs should be integrated into control or safety logic where appropriate, and their functionality should be verified during commissioning and proof testing.


## 7.6 Importance of a Lifecycle Perspective

Commissioning, calibration, proof testing, and diagnostics should not be treated as isolated activities. Together, they form a measurement lifecycle framework that ensures oxygen analyzers remain accurate, reliable, and fit for purpose over time.

Following recognized standards—particularly ASTM practices for analyzer validation—provides a consistent and transparent basis for decision-making, supports regulatory and safety requirements, and reinforces confidence in oxygen measurement as a critical process variable.

## 8. MOD-1040 Optical Oxygen Analyzer

The MOD-1040 Process Oxygen Analyzer is a high-performance, in-situ measurement solution engineered for accurate and reliable oxygen analysis directly in high-pressure gas pipelines. Designed for safety-critical and mission-critical applications, the MOD-1040 eliminates the need for gas sample extraction, conditioning systems, or pressure reduction—fundamentally changing how oxygen is measured in hazardous and high-integrity process environments.



*Fig. 16 – MOD-1040 Optical Oxygen Analyzer*

By combining advanced photonics-based sensing with rugged, explosion-proof construction, the MOD-1040 delivers precise oxygen measurements across the full concentration range—from trace ppm levels to 100% O<sub>2</sub>—even under extreme pressure, temperature, and environmental conditions. This makes the MOD-1040 particularly well suited for hydrogen production, hydrogen compression and storage, natural gas and refinery processes, chemical production units, and industrial gas systems, where oxygen ingress represents a critical safety, quality and reliability risk.

### 8.1 Advanced Photonics-Based Measurement Principle

The MOD-1040 is based on advanced photonics technology utilizing luminescence quenching of a specially designed sensor dye. The sensing dye is immobilized on a robust support foil, forming a stable and highly selective sensing layer.




Fig. 17 – Luminescence quenching principle

This optical principle is inherently stable, drift-free, and immune to many of the limitations associated with electrochemical, paramagnetic, zirconia or TDL-based oxygen analyzers—especially under high pressure.

- The sensing layer is stimulated with red light emitted from an integrated optical source.
- The dye responds by emitting luminescence in the near-infrared (NIR) region of the electromagnetic spectrum.
- When molecular oxygen is present, it interacts with the excited dye molecules and quenches the luminescence.
- This quenching causes a reversible change in both luminescence intensity and lifetime, which is directly proportional to the oxygen concentration.
- The MOD-1040 precisely analyzes these optical changes to determine the true oxygen content in real time.

## 8.2 Field-Replaceable Optical Sensor Architecture

The sensor technology is implemented as a compact, plug-in optical module, specifically designed for industrial field conditions:

- No electrolyte consumption
- No membrane degradation
- Easily replaceable sensor spot
- No frequent recalibration cycles

An integrated infrared light source excites the sensing layer, while the emitted near-infrared luminescence is detected and analyzed internally. The fully reversible quenching mechanism ensures long-term stability and consistent performance across a wide range of operating conditions.

### **8.3 True In-Situ Oxygen Measurement—Without Sample Extraction**

Industrial gas and hydrogen production systems often operate in high-pressure, hazardous zones, where minimizing leak paths is a primary safety objective. Traditional oxygen analysis technologies are not designed to withstand such conditions and therefore rely on sample extraction, pressure reduction, and conditioning systems—introducing additional failure points, maintenance burden, and safety risks.

The MOD-1040 overcomes these limitations by enabling direct, in-situ oxygen measurement inside the process pipeline, even at very high pressures. This capability:

- Eliminates sample systems entirely
- Reduces potential leak points
- Simplifies hazardous area classification
- Lowers installation and lifecycle costs
- Improves measurement response time and reliability

As a result, MOD-1040 allows engineers to confidently deploy oxygen measurement even in locations previously considered impractical or unsafe.

### **8.4 Engineered for Hydrogen and Safety-Critical Applications**

In green and blue hydrogen production, oxygen contamination can lead to:

- Increased explosion risk
- Catalyst degradation
- Reduced product purity
- Accelerated material aging

The MOD-1040 is specifically engineered to address these challenges. Its optical sensing technology is ideally suited for hydrogen service, offering fast response, high selectivity, and intrinsic safety, even under pressures reaching hundreds of bar.

Automatic pressure and temperature compensation ensures that the reported oxygen concentration reflects true process conditions, not measurement artifacts caused by fluctuating operating parameters.



## 8.5 Key Capabilities and Technical Specifications

### Measurement Performance:

- In-situ oxygen measurement range: 1 ppm to 100% O<sub>2</sub>
- Response time (T90): < 5 seconds
- Automatic pressure and temperature compensation

### Operating Conditions:

- Maximum operating pressure: up to 350 barg
- Ambient temperature range: -10 to +60 °C
- Designed for continuous operation in harsh industrial environments

### Safety and Compliance:

- Explosion-proof certification: II 2 G Ex db IIC T4 Gb
- SIL-2 certified in accordance with IEC 61508-2:2010
- Suitable for ATEX / IECEx Zone 1 installations

### Integration and Connectivity:

- Modbus communication for DCS, PLC and HMI systems
- Bluetooth interface for commissioning, diagnostics and maintenance
- Industrial-standard 4–20 mA analog inputs and outputs, enabling seamless integration with pressure and temperature transmitters and direct communication with DCS, PLC and safety shutdown systems.
- Compact, rugged design for direct pipe mounting

## 8.6 A New Standard for Oxygen Analysis in High-Pressure Processes

The MOD-1040 Process Oxygen Analyzer is not simply another analyzer option—it represents a step change in how oxygen is measured in high-pressure and hazardous environments. By combining photonics-based luminescence quenching, in-situ installation, and uncompromising safety certifications, the MOD-1040 enables safer plant designs, higher measurement integrity, and lower total cost of ownership.

For hydrogen production, refinery operations, and advanced industrial gas systems, the MOD-1040 sets a new benchmark for precision, safety, and operational efficiency

## 8.7 Calibration, Validation and Installation

Accurate oxygen measurement in industrial processes is not solely a function of sensor technology; it is equally governed by calibration philosophy, installation quality, and ongoing validation practices. In safety-critical and asset-integrity-driven applications, oxygen analyzers must be treated as metrological instruments embedded within a broader process system, rather than as standalone devices.



### *Calibration Strategy and Traceability*

For in-situ oxygen analyzers, calibration philosophy differs fundamentally from extractive systems. Because the measurement is performed directly at process conditions—pressure, temperature, and composition—calibration gases do not need to replicate the full process matrix, but must be traceable, stable, and applied in a controlled manner.

Best practice includes:

- Zero verification using oxygen-free or oxygen-depleted gases (e.g., nitrogen or hydrogen-rich purge streams).
- Span verification using certified oxygen mixtures selected to bracket the relevant operating range rather than full-scale values.
- Documentation of as-found and as-left values, enabling trend-based assessment of sensor drift rather than periodic recalibration alone.

Optical oxygen measurement technologies, in particular, benefit from calibration approaches that decouple sensor physics from gas transport effects, allowing longer calibration intervals and reducing dependency on frequent manual intervention.

### *Validation and Functional Testing*

Validation extends beyond calibration accuracy and must address whether the analyzer continues to perform its intended function under real operating conditions. This is especially critical in trace-oxygen and safety-related applications.

Recommended validation practices include:

- Verification of response time under representative process dynamics.
- Monitoring of internal diagnostics such as optical signal strength, excitation source health, and temperature stability.
- Periodic challenge tests aligned with the defined proof-test interval for safety instrumented functions (SIFs), where applicable.

Validation should confirm not only measurement correctness but also fault detectability, ensuring that internal failures are identified before they compromise process safety or asset integrity.

### *Installation and Process Interface Design*

Installation quality often dominates real-world performance, particularly for trace oxygen applications. In-situ analyzers eliminate transport delay and many leak paths associated with extractive systems, but they require careful attention to the process interface.

Key installation considerations include:

- Selection of locations with representative flow and composition, avoiding stagnant zones and boundary layers.
- Mechanical design that prevents air ingress during maintenance, pressure cycling, or thermal contraction.



- Control of surface finishes and materials to minimize oxygen adsorption and desorption, which can distort low-ppm measurements.

When properly installed, in-situ analyzers provide a more faithful representation of process oxygen activity than downstream sampling systems, particularly in high-pressure and hydrogen-rich environments.

#### *Balance-of-Plant (BoP) Perspective: Oxygen, Hydrogen, and Intelligence*

Modern hydrogen and energy systems increasingly rely on integrated Balance-of-Plant architectures, where multiple analyzers and sensors feed higher-level optimization and safety logic. Within this context, oxygen measurement plays a pivotal role.

A coordinated BoP approach typically integrates:

- Oxygen analysis for inerting, safety, and corrosion control.
- Hydrogen analysis for purity, leak detection, and process efficiency.
- Advanced analytics or AI-based models that correlate sensor data with operational outcomes.

In such architectures, oxygen is often the leading safety variable—its presence defines ignition risk, material compatibility, and allowable operating envelopes. High-integrity, in-situ oxygen measurement therefore becomes a cornerstone of system-level intelligence, enabling faster response, reduced uncertainty, and more resilient operation.

## 9. Applications of Oxygen Measurement in Industrial Processes

Oxygen measurement is a foundational safety, quality, and reliability function across a wide range of process industries. While the physical measurement principle may be similar, the role oxygen plays, the acceptable limits, and the consequences of failure vary significantly by application. This chapter reviews key industrial use cases, emphasizing how process conditions, safety objectives, and system design requirements shape oxygen analyzer selection and deployment.

### 9.1 Natural Gas Processing and Transmission

#### *Role of Oxygen Measurement*

In natural gas production, processing, and transmission systems, oxygen is considered a contaminant. Even trace concentrations can introduce multiple risks:

- Corrosion acceleration, particularly in the presence of moisture and acid gases (CO<sub>2</sub>, H<sub>2</sub>S)
- Degradation of pipeline materials and compressor components
- Increased explosion risk during maintenance, purging, or upset conditions
- Catalyst poisoning in downstream processing units

Oxygen may enter natural gas systems through:

- Air ingress during maintenance or pigging
- Leaks in compressor seals or sampling systems
- Inadequate purging during startup or shutdown
- Interface with atmospheric storage or blending operations

#### *Measurement Challenges*

Natural gas environments are technically demanding:

- High pressures (often >100 barg)
- Variable composition (methane, ethane, heavier hydrocarbons)
- Presence of CO<sub>2</sub>, H<sub>2</sub>S, and water vapor
- Risk of adsorption/desorption effects at trace oxygen levels

In such systems, sample system integrity often dominates overall measurement uncertainty, particularly when oxygen limits are in the low ppmv range.

#### *Implementation Considerations*

- Oxygen is typically monitored continuously at critical points such as compressor discharge, dehydration outlets, or custody-transfer interfaces.
- Measurement ranges often span from ppmv to low % v/v, depending on system function.
- Analyzer installations must comply with hazardous-area requirements (Zone 1 / Class I Div 1).



- Diagnostics are essential to distinguish true oxygen ingress from analyzer or sampling faults.

Optical in-situ or high-integrity extractive analyzers (e.g., MOD-1040 used as an example of this class) are often selected because they minimize sample handling and reduce air ingress risk.

## 9.2 Hydrogen Production, Processing, and Storage

### *Safety Context*

Hydrogen presents one of the most demanding safety environments for oxygen measurement. Its:

- Extremely low ignition energy
- Wide flammability range
- High diffusivity

mean that even small oxygen concentrations can create hazardous mixtures.

Hydrogen is produced through multiple pathways:

- Steam Methane Reforming (SMR)
- Water electrolysis (alkaline, PEM, SOEC)
- Biomass gasification
- Reforming of renewable feedstocks

Across all routes, oxygen control is essential to maintain concentrations well below the Limiting Oxygen Concentration (LOC).

### *Measurement Objectives*

Oxygen monitoring in hydrogen systems serves to:

- Prevent formation of flammable or detonable mixtures
- Verify inerting effectiveness
- Detect air ingress in high-pressure systems
- Support safe startup, shutdown, and maintenance procedures

### *Applications include:*

- Hydrogen dryers and purifiers
- High-pressure hydrogen compressors
- Storage vessels and tube trailers
- Distribution pipelines
- Fuel cell supply systems

### *Technical Challenges*

- Pressures commonly exceed 200 barg
- Hydrogen's low molecular weight complicates extractive sampling
- Traditional electrochemical sensors may be unsuitable due to cross-sensitivity or lifetime limitations



- Fail-safe behavior is critical: fail-low can be dangerous if oxygen ingress goes undetected

#### *System Design Implications*

- Analyzer selection must consider failure direction, diagnostic coverage, and proof-test strategy.
- In many hydrogen safety functions, analyzers are part of a Safety Instrumented Function (SIF) rather than purely operational monitoring.
- Technologies that do not consume oxygen and that operate without pressure reduction are often preferred.

In this context, advanced optical analyzers (such as MOD-1040-type systems) are frequently deployed as part of integrated hydrogen safety architectures.

### 9.3 Chemical and Petrochemical Processing

#### *Process Role of Oxygen*

In chemical and petrochemical plants, oxygen measurement supports:

- Explosion prevention in reactors and storage vessels
- Inerting verification during batch operations
- Protection of flammable solvent systems
- Compliance with safety and environmental regulations

Oxygen limits are often defined by:

- Material safety data
- Process hazard analysis (PHA)
- Layer of Protection Analysis (LOPA)
- Insurance and regulatory requirements

#### *Typical Applications*

- Reactor headspace monitoring
- Nitrogen inerting systems
- Solvent storage tanks
- Polymerization units
- Distillation columns handling flammable feeds

#### *Measurement Requirements*

- Rapid response during transient events
- Reliable operation across wide temperature ranges
- Resistance to solvent vapors and process contaminants
- Integration with control and safety systems

In many chemical applications, oxygen analyzers are used in dual roles:

- As control inputs for inert gas flow
- As independent safeguards triggering alarms or trips



The ability to clearly diagnose sensor health and distinguish process upsets from analyzer faults is essential.

## 9.4 High-Pressure Gas Production and Processing

### *Importance of Pressure Capability*

Many modern gas-processing applications operate at elevated pressures to:

- Improve energy efficiency
- Reduce equipment size
- Enable downstream compression or storage

However, pressure significantly influences:

- Oxygen partial pressure
- Gas density and diffusion behavior
- Mechanical stress on analyzer components

### *Design Implications*

- Pressure reduction prior to measurement introduces:
  - Transport delay
  - Additional leak paths
  - Potential oxygen enrichment in trapped volumes
- High-pressure-capable analyzers allow:
  - Direct measurement at process pressure
  - Improved safety integrity
  - Faster response times

Systems designed for operation up to 200 barg and beyond (as exemplified by MOD-1040-class analyzers) address these challenges by eliminating pressure letdown and minimizing sample-system complexity.

## 9.5 Cross-Application Considerations

Regardless of industry, several principles apply universally:

- Oxygen analyzers must be evaluated as complete systems, not standalone sensors.
- Trace oxygen applications are often limited by:
  - Air ingress
  - Adsorption/desorption
  - Sample transport delay
- Safety relevance depends on:
  - Failure mode (fail-high vs fail-low)
  - Diagnostic capability
  - Integration with safety logic



- Technology selection should follow the Safety Requirement Specification (SRS), not vice versa.

## 10. Glossary and Terms

### **Accuracy**

The closeness of agreement between a measured oxygen value and the true or reference value, typically expressed as a percentage of reading or span.

### **Adsorption / Desorption**

Surface phenomena in which oxygen molecules temporarily adhere to (adsorption) or are released from (desorption) internal surfaces of sample lines, regulators, and fittings, often dominating bias and response time in trace oxygen systems.

### **Air Ingress**

Unintended entry of ambient air into a process or sample system through leaks, permeation, or pressure imbalance, representing a primary error source for low-ppmv oxygen measurements.

### **Analyzer System**

The complete measurement assembly including the sensor, sample probe, sample conditioning, transport lines, analyzer electronics, utilities, and interfaces—evaluated as a whole rather than as a sensor alone.

### **As-Found / As-Left**

Documentation of analyzer performance before (as-found) and after (as-left) calibration or maintenance, essential for traceability and safety integrity verification.

### **Calibration**

Adjustment of analyzer output to match known reference points, typically zero and span, using traceable calibration gases or solutions.

### **Common Cause Failure (CCF)**

A failure mechanism that simultaneously affects multiple redundant channels due to a shared cause, such as environmental stress, design flaw, or maintenance error.

### **Diagnostic Coverage (DC)**

The fraction of dangerous failures detected by internal diagnostics, expressed as a percentage and used in Safety Integrity Level (SIL) calculations.

### **Dissolved Oxygen (DO)**

Oxygen present in liquid media, commonly expressed in mg/L or ppb, governed by solubility, temperature, pressure, and mass-transfer kinetics.

### **Fail-High / Fail-Low**

Defined analyzer behavior under fault conditions:

- *Fail-high*: output drives to a higher oxygen indication.



- *Fail-low*: output drives to a lower oxygen indication.  
The dangerous direction depends on the Safety Instrumented Function (SIF).

**Gas Chromatography (GC)**

A separation-based analytical technique that quantifies oxygen as one component of a gas mixture, offering high selectivity but slower response and higher system complexity.

**Hazardous Area Classification**

Designation of areas where flammable gases or vapors may be present, determining requirements for explosion protection (e.g., ATEX, IECEx).

**In-Situ Analyzer**

An analyzer measuring oxygen directly in the process without sample extraction, minimizing transport delay but requiring robust process interfaces and fouling control.

**Limit of Detection (LOD)**

The lowest oxygen concentration distinguishable from background noise with a defined statistical confidence.

**Mass Spectrometry (MS)**

An analytical technique measuring ionized species based on mass-to-charge ratio, providing multi-component analysis with high sensitivity but significant complexity.

**Measurand**

The specific quantity intended to be measured, such as oxygen volume fraction, partial pressure, or dissolved concentration.

**Optical (Luminescence / Fluorescence) Oxygen Measurement**

A technique based on oxygen-dependent quenching of luminescent materials, offering drift-free behavior, fast response, and minimal consumables.

**Paramagnetic Oxygen Measurement**

A method exploiting the strong paramagnetism of oxygen molecules relative to other gases, commonly used for percent-level measurements.

**Partial Pressure**

The pressure contribution of oxygen in a gas mixture, equal to the total pressure multiplied by the oxygen mole fraction.

**Proof Test Interval (PTI)**

The maximum allowable time between functional tests intended to reveal dangerous undetected failures in safety-related analyzers.

**Response Time (T90 / T95)**

The time required for an analyzer to reach 90% or 95% of a final value following a step change in oxygen concentration.

**Safety Integrity Level (SIL)**

A discrete level (SIL 1–4) specifying the required risk reduction for a Safety Instrumented Function, as defined in IEC 61508 / IEC 61511.

**Safety Instrumented Function (SIF)**

A safety function implemented by sensors, logic solvers, and final elements to reduce process risk to a tolerable level.

**Sample Conditioning**

Processes applied to extracted samples—such as filtration, pressure reduction, temperature control, and drying—to ensure compatibility with the analyzer.

**Span Gas**

A calibration gas with a known oxygen concentration near the upper end of the measurement range.

**Trace Oxygen**

Oxygen measurement typically below 100 ppmv, where contamination, adsorption, and system integrity dominate measurement uncertainty.

**Zero Gas**

A calibration gas with negligible oxygen content, used to establish the analyzer zero reference.